

Danish Defence Acquisition And Logistics Organization

The Danish Defence Acquisition and Logistics Organization Tender for Energy Dispersive Spectroscopy (EDS) detector

Question submitted	Answer	Date of answering
Question no. 1:		
Typically on a Zeiss EVO LS25 you will find 10 mm2, 30 mm2 or in rare cases perhaps maximum 60 mm2 detectors installed. Why is an expensive super-large (80 mm2 or more) detector required? Is it a minus if a 60 mm2 detector is offered? If yes, is there a bonus if a 100 mm2 detector is offered?	A detector with an area of at least 80 mm ² was specified based on our need for a detector capable of performing gun shot residue analysis and particle analysis within a reasonable timeframe.	11-SEP-2018
	An offer for a 60 m ^{m2} detector is not conditional and will be discarded.	
	An offer for a 100 mm² is conditional and will be evaluated based on the same criteria as any other offer. The criteria is price.	
Question no. 2:		
Could you elaborate why a motorized retraction system for the detector is needed?	A motorized retraction system is needed when large and irregular objects is introduced into the sample chamber and the EDS detector needs to be retracted to prevent any collision damage.	11-SEP-2018
Question no. 3:		
The tender requires software for automated particle analysis. Yet no software is required to control the SEM stage. Could you please comment?	For the purpose of this tender "automated particle analysis" is understood as the ability of a software to recognize individual particles, estimate their size and chemical class. The particle analysis only needs to be done an a single static frame, thus software control of the SEM stage is not a requirement.	11-SEP-2018
Question no. 4:		
The tender requires: "The detector window shall be opaque to light". However, no detector on the market can comply with that unless a beryllium window is offered, which conflicts with the requirement to be able to analyze down to Carbon.	The requirement: "The detector window shall be opaque to light" shall be understood as the: "The detector window shall be opaque to the small amount of light generated through cathodoluminescence.	11-SEP-2018

DALO Version 4.0

() All detectors are opaque to the small amount of light generated through cathodoluminescence, but no detector is opaque to e.g. daylight. Could you please comment?		
Question no. 5: In the tender conditions is written: "Similar products, able to fulfill the same requirements / needs and of an equivalent quality, can therefore also be offered". Does this apply to any "shall" requirement in the tender	Yes, but the tenderer must prove that the products are comparable.	11-SEP-2018
Question no. 6: Will a 60mm2 detector area be acceptable	Requirement #12 An offer for a 60 mm² detector is not conditional and the offer will be discarded.	11-SEP-2018